



**CAB/ILAC-TP/092/INF
2008-10-17**

CALIBRATION SERVICES PURCHASE ORDER CRITERIA

Calibration vendors shall meet the following requirements:

- Maintain a current ISO/IEC 17025 accreditation through an ILAC Full Member body for the scope of the calibration service being requested or be subject to an assessment by CSA International staff to verify compliance with ISO/IEC 17025 for the service provided.
- Provide information on the calibration certificate/report detailed in ISO/17025:2005 Clauses 5.10.2 and 5.10.4.
- Provide statement of uncertainties, including:
 1. individual uncertainty values for each calibration measurement, or
 2. total calculated measurement uncertainty and a statement that individual results are available upon request, or
 3. statement of conformance to a tolerance (eg.: the uncertainty in the measurement system did not exceed 25% (4:1 test uncertainty ratio) of the acceptable tolerance for each characteristic calibrated, unless otherwise noted in the report) and a statement that individual results are available upon request.